

# CERTIFICATE OF CALIBRATION ISSUED BY SERVICECAL LTD

DATE OF ISSUE 12th April 2023

CERTIFICATE NUMBER

DATE OF RECEIPT 6th April 2023

U367569



0152



UKAS Accredited Laboratory (0152)  
Unit A9 Axis Point  
Hareshill Business Park, Hilltop Road  
Heywood, Lancs  
OL10 2RQ

Tel 01706 367008

Fax 01706 622469

EEmail [info@servicecal.co.uk](mailto:info@servicecal.co.uk)

Web [www.servicecal.co.uk](http://www.servicecal.co.uk)

Page 1 of 4

Approved By **Martin Ashworth**

Signatory

INSTRUMENT DESCRIPTION	Digital Multimeter	✓
MANUFACTURER	Agilent	
MODEL NUMBER	34401A	
SERIAL NUMBER	MY44002035	✓
CUSTOMER REFERENCE NUMBER		
CUSTOMER	Gjc Instruments Ltd	
ADDRESS	North West House, Near Brook Hall, Off Chester Road, Tattenhall, Cheshire, CH3 9AH	
ORDER NUMBER	2023LAB002	
REMARKS	The instrument was externally cleaned prior to calibration	

The reported expanded uncertainty is based on a standard uncertainty multiplied by a coverage factor  $k=2$ , providing a level of confidence of approximately 95%. The uncertainty evaluation has been carried out in accordance with UKAS requirements.

This certificate is issued in accordance with the laboratory accreditation requirements of the United Kingdom Accreditation Service, and is subject to our standard terms and conditions. The certificate and results apply only to the item above that has been calibrated.

It provides traceability of measurement to recognised National Standards, and to units realised at the National Physical Laboratory or other recognised National Standards Laboratories. This certificate may not be reproduced other than in full,

except with the prior written approval of the issuing laboratory - Servicecal Ltd.

✓ checked @  
GJC today  
12/4/23

**CUSTOMER** GJC Instruments Ltd  
**ADDRESS** Northwest House, Tattenhall, Cheshire. CH3 9AH.  
**JOB NUMBER** 367569  
**DESCRIPTION** Digital Multimeter  
**MANUFACTURER** Agilent  
**MODEL NUMBER** 34401A  
**SERIAL NUMBER** MY44002035  
**CUSTOMER REFERENCE** None  
**PREVIOUS CERTIFICATE** U347333 **ISSUED BY** 0152

**CALIBRATION INFORMATION.**

The instrument was calibrated against the laboratory's standards which are traceable to UK National Standards or are derived by approved ratio techniques. The ambient temperature and relative humidity during the tests were  $20^{\circ}\text{C} \pm 2^{\circ}\text{C}$  and  $50\% \pm 20\%$  RH respectively. The uncertainty limits quoted refer to the measured values only, with no account being taken of the instrument's ability to maintain its calibration.

The limits have been derived in accordance with UKAS requirements and are for a confidence level of approximately 95% using a coverage factor of  $k=2$ .

The instrument was allowed to stabilise in the laboratory for a minimum of 4 hours prior to measurement being performed.

**SPECIFICATION USED OR EXTENT OF CALIBRATION PERFORMED.**

The instrument was calibrated using laboratory measurement procedures:  
UKASP001,002,003,005,007,008,009,053

Checking of instrument ranges to manufacturer's specification  
Specification taken from Agilent 34401A users guide p/n 34401-90004 March 2000 edition 5.

**SUMMARY**

The calibration results are recorded as found upon receipt of the instrument.

At the points tested the instrument was found to meet specification, due allowance having been made for the uncertainty of measurement.

The Decision Rule : ILAC-G8:09/2019 4.2.3 Non Binary Statement with Guard Band has been applied.  
See Appendix A attached.

The instrument was externally cleaned prior to calibration.

Measured Mains Voltage = 235 V.

**CALIBRATED BY** M.Ashworth **Date** 12th April 2023

**CHECKED BY** M.Ashworth **Date** 12th April 2023

✓ checked @  
GJC  
Aday  
12/04/23

CALIBRATION RESULTS

VOLTAGE RANGE	APPLIED FREQUENCY	APPLIED VOLTAGE	INSTRUMENT READING
100 mV	DC	100.0000 mV	100.0015 mV
100 mV	DC	-100.0000 mV	-100.0017 mV
1 V	DC	1.000000 V	1.000021 V
1 V	DC	-1.000000 V	-1.000020 V
10 V	DC	1.000000 V	01.00001 V
10 V	DC	3.000000 V	03.00004 V
10 V	DC	6.000000 V	06.00007 V
10 V	DC	10.00000 V	10.00012 V
10 V	DC	11.00000 V	11.00014 V
10 V	DC	-1.000000 V	-01.00002 V
10 V	DC	-3.000000 V	-03.00004 V
10 V	DC	-6.000000 V	-06.00007 V
10 V	DC	-10.00000 V	-10.00011 V
10 V	DC	-11.00000 V	-11.00014 V
100 V	DC	100.0000 V	100.0005 V
100 V	DC	-100.0000 V	-100.0005 V
1000 V	DC	1000.0000 V	0999.993 V
1000 V	DC	-1000.0000 V	-0999.992 V
100 mV	500 Hz	100.0000 mV	099.9757 mV
1 V	500 Hz	1.000000 V	0.999759 V
10 V	500 Hz	10.00000 V	09.99768 V
100 V	500 Hz	100.0000 V	099.9471 V
750 V	500 Hz	500.0000 V	0499.723 V
100 mV	1 kHz	100.0000 mV	099.9753 mV
1 V	1 kHz	1.000000 V	0.999775 V
10 V	1 kHz	1.000000 V	00.99976 V
10 V	1 kHz	5.000000 V	04.99876 V
10 V	1 kHz	10.00000 V	09.99783 V
100 V	1 kHz	100.0000 V	099.9519 V
750 V	1 kHz	500.0000 V	0499.750 V
100 mV	10 kHz	100.0000 mV	099.9747 mV
1 V	10 kHz	1.000000 V	0.999745 V
10 V	10 kHz	10.00000 V	09.99788 V
100 V	10 kHz	100.0000 V	099.9510 V
100 mV	50 kHz	100.0000 mV	099.9750 mV
1 V	50 kHz	1.000000 V	0.999643 V
10 V	50 kHz	10.00000 V	09.99932 V
100 V	50 kHz	100.0000 V	099.9467 V

The reported expanded uncertainty is based on a standard uncertainty multiplied by a coverage factor  $k=2$ , providing a level of confidence of approximately 95%, the uncertainty evaluation has been carried out with UKAS requirements.

CALIBRATED BY M.Ashworth Date 12th April 2023

CHECKED BY M.Ashworth M Ashworth Date 12th April 2023

✓ checked @  
GJC  
today 12/4/23

CURRENT RANGE	APPLIED FREQUENCY	APPLIED CURRENT	INSTRUMENT READING
10 mA	DC	10.00000 mA	09.99992 mA
100 mA	DC	100.0000 mA	099.9983 mA
1 A	DC	1.000000 A	0.999671 A
3 A	DC	1.900000 A	01.89937 A
3 A	1 kHz	1.000000 A	00.99942 A
1 A	1 kHz	1.000000 A	0.999451 A

FRONT INPUT TERMINALS

RESISTANCE RANGE	APPLIED RESISTANCE	4-WIRE READING
100 Ω	99.99799 Ω	100.0006 Ω
1 kΩ	1.0000006 kΩ	1.000037 kΩ
10 kΩ	1.0000006 kΩ	01.00003 kΩ
10 kΩ	9.999982 kΩ	10.00020 kΩ
100 kΩ	100.00122 kΩ	100.0044 kΩ
1 MΩ	0.999965 MΩ	1.000017 MΩ
10 MΩ	9.998718 MΩ	09.99776 MΩ

REAR INPUT TERMINALS

RESISTANCE RANGE	APPLIED RESISTANCE	4-WIRE READING
100 Ω	99.99799 Ω	100.0004 Ω
1 kΩ	1.0000006 kΩ	1.000038 kΩ
10 kΩ	9.999982 kΩ	10.00020 kΩ
100 kΩ	100.00122 kΩ	100.0047 kΩ
1 MΩ	0.9999648 MΩ	1.000018 MΩ
10 MΩ	9.998718 MΩ	09.99772 MΩ

UNCERTAINTY OF MEASUREMENT:

DC Current	10 mA ± 50 ppm	100 mA ± 60 ppm
	1 A-2 A ± 100 ppm	
Resistance	100 Ω, 1 kΩ, 10 kΩ, 100 kΩ	± 10 ppm
	1 MΩ ± 10 ppm	10 MΩ ± 30 ppm
DC Voltage	100 mV - 1000 V	± 7 ppm
AC Current	1 A (1kHz)	± 0.014%
AC Voltage	100 mV to 100 V (500 Hz, 1, 10, 50 kHz)	± 0.015%
AC Voltage	500V (500 Hz, 1 kHz)	± 0.030% + 50 μV

1 LSD must be added to the above uncertainties to allow for the resolution of the instrument under test.

The reported expanded uncertainty is based on a standard uncertainty multiplied by a coverage factor  $k=2$ , providing a level of confidence of approximately 95%, the uncertainty evaluation has been carried out with UKAS requirements.

CALIBRATED BY M.Ashworth

Date 12th April 2023

CHECKED BY

M.Ashworth

Date 12th April 2023

checked @  
GTC today